

Laboratory Bottles with Thread, DURAN®, DURAN Group / IDL

Laboratory bottles with graduation, ISO 4796-1, with Retrace Code (Batch Certificate of Performance via Internet). Clear glass according to USP/EP and JP.

① Without cap and pouring ring

Volume ml	Thread GL	O.D. mm	Height mm	Order No.	PU pc/Pack
10 ²⁾	25	36	50	10 21 10108	10
25 ²⁾	25	36	70	11 21 80114	10
50	32	46	87	11 21 80117	10
100 ¹⁾	45	56	100	11 21 80124	10
150 ¹⁾	45	62	110	10 21 10129	10
250 ¹⁾	45	70	138	11 21 80136	10
500 ¹⁾	45	86	176	11 21 80144	10
750 ¹⁾	45	95	203	10 21 10153	10
1000 ¹⁾	45	101	225	11 21 80154	10
2000 ¹⁾	45	136	260	11 21 80163	10
3500	45	160	295	10 21 10162	1
5000	45	182	330	11 21 80173	1
10000	45	227	410	11 21 80186	1
15000	45	268	445	11 21 80188	1
20000	45	288	505	11 21 80191	1



WWEQ2034

② With screw cap and pouring ring made of PP (blue)

Volume ml	Thread GL	O.D. mm	Height mm	Order No.	PU pc/Pack
10 ²⁾	25	36	50	10 21 15108	10
25 ²⁾	25	36	74	11 21 85114	10
50	32	46	91	11 21 85117	10
100 ¹⁾	45	56	105	11 21 85124	10
150 ¹⁾	45	62	114	10 21 11129	10
250 ¹⁾	45	70	143	11 21 85136	10
500 ¹⁾	45	86	181	11 21 85144	10
750 ¹⁾	45	95	207	10 21 11153	10
1000 ¹⁾	45	101	230	11 21 85154	10
2000 ¹⁾	45	136	265	11 21 85163	10
3500	45	160	299	10 21 11162	1
5000	45	182	335	11 21 85173	1
10000	45	227	415	11 21 85186	1
15000	45	268	450	11 21 85188	1
20000	45	288	510	11 21 85191	1



1) A bead is molded in the glass on the shoulder of the 100 to 2000 ml size to indicate the nominal capacity level.

2) With a specially formed glass rim for better pouring (hence no need for an additional pouring ring made of plastic).

Important: Autoclave these flasks only with a loosely screwed cap (do not close tightly).

