

Laboratory Bottles with Thread, Amber, DURAN®, DURAN Group / IDL



Laboratory bottles with graduation, without cap, ISO 4796-1, with Retrace Code (Batch Certificate of Performance via Internet). Clear glass 1 according to USP/EP/JP. Light shield up to a light wave length of up to 500 nm. External application of the paint ensures unchanged DURAN® characteristics in the interior of the bottle.

Vol. ml	Thread GL	O.D. mm	Height mm	Order No.	PU pc/Pack
10 ²⁾	25	36	50	10 21 10608	10
10 ³⁾²⁾	25	36	50	10 21 10609	10
25 ²⁾	25	36	70	11 21 80614	10
50	32	46	87	11 21 80617	10
100 ¹⁾	45	56	100	11 21 80624	10
150 ¹⁾	45	62	110	10 21 10629	10
250 ¹⁾	45	70	138	11 21 80636	10
500 ¹⁾	45	86	176	11 21 80644	10
750 ¹⁾	45	95	203	10 21 10651	10
1000 ¹⁾	45	101	225	11 21 80654	10
2000 ¹⁾	45	136	260	11 21 80663	10
3500	45	160	295	10 21 10669	1
5000	45	182	330	11 21 80673	1
10000	45	227	410	11 21 80686	1
20000	45	288	505	11 21 80691	1

¹⁾ A bead is molded in glass on the shoulder of the 100 to 2000 ml size to indicate the nominal capacity level.

²⁾ With a specially formed glass rim for better pouring (hence no need for an additional pouring ring made of plastic).

³⁾ Complete

Important: Autoclave these flasks only with a loosely screwed cap (do not close tightly).

Note: This item is for example also available with cap.

