

Volumetric Flasks, Amber Glass, Class A, BRAND



- ① **Volumetric flasks, BLAUBRAND®, Class A, amber glass, white graduation, with PP stopper**
DE-M marked, with batch number and enclosed batch certificate.
Boro 3.3, DIN EN ISO 1042. Calibrated to contain (In). Ring marks and inscriptions in high-contrast white enamel. With PP stopper. Original packs of 2, including batch certificate. On request, also available with individual certificate or DAkkS calibration certificate.

Volume ml	Error limit ± ml	NS	Order No.	PU pc/Pack
5 W	0.04	10/19	14 49 37450	2
10 W	0.04	10/19	14 49 37460	2
20	0.06	12/21	10 49 37412	2
20	0.04	10/19	14 49 37464	2
25	0.06	10/19	10 49 37413	2
25	0.04	10/19	14 49 37469	2
50	0.06	12/21	14 49 37475	2
50 W	0.10	14/23	14 49 37476	2
100	0.10	14/23	14 49 37480	2
200	0.15	14/23	14 49 37482	2
250	0.15	14/23	14 49 37486	2
500	0.25	19/26	14 49 37490	2
1000	0.4	24/29	14 49 37494	1

W = wide neck

- ② **Volumetric flasks, BLAUBRAND®, Class A, amber glass, white graduation, glass stoppers**
DE-M marked, Boro 3.3, DIN EN ISO 1042. Calibrated to contain (In). Ring marks and inscriptions in high-contrast white enamel. With glass stoppers. All BLAUBRAND® volumetric flasks are with batch number and enclosed batch certificate. On request, also available with individual certificate or DAkkS calibration certificate.

Volume ml	Error limit ± ml	NS	Order No.	PU pc/Pack
5 W	0.04	10/19	10 49 37438	2
10 W	0.04	10/19	10 49 37443	2
20	0.04	10/19	10 49 37446	2
25	0.04	10/19	10 49 37447	2
50	0.06	12/21	10 49 37448	2
50 W	0.08	14/23	10 49 37445	2
100	0.10	14/23	10 49 37449	2
200	0.15	14/23	10 49 37450	2
250	0.15	14/23	10 49 37451	2
500	0.25	19/26	10 49 37452	2
1000	0.4	24/29	10 49 37453	2

W = wide neck

